			M	
	Application No.	Ap	plicant(s)	
	10/777,249	NE	SS, GREGORY O.	
Notice of Allowability	Examiner		t Unit	
	Philip R. Smith	37:	39	
The MAILING DATE of this communication ap All claims being allowable, PROSECUTION ON THE MERITS I herewith (or previously mailed), a Notice of Allowance (PTOL-8 NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT of the Office or upon petition by the applicant. See 37 CFR 1.3	S (OR REMAINS) CLOSE( 5) or other appropriate com RIGHTS. This application	D in this application will	tion. If not included be mailed in due course. TH	
1. $\boxtimes$ This communication is responsive to <u>the correspondence</u>	e of 9/18/2006.	<u>.</u> :		
2. The allowed claim(s) is/are 2-10 and 31-50.	:	:		
<ul> <li>3. Acknowledgment is made of a claim for foreign priority</li> <li>a) All b) Some* c) None of the:</li> <li>1. Certified copies of the priority documents hat</li> <li>2. Certified copies of the priority documents had</li> </ul>	ve been received.			:
3. Copies of the certified copies of the priority of			onal stage application from the	ne .
International Bureau (PCT Rule 17.2(a)).	goodinomo navo boon roco		mar dago application from a	.0
* Certified copies not received:		1		
Applicant has THREE MONTHS FROM THE "MAILING DATE noted below. Failure to timely comply will result in ABANDON THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		file a reply com	plying with the requirements	
4. A SUBSTITUTE OATH OR DECLARATION must be sub INFORMAL PATENT APPLICATION (PTO-152) which g				:
5. CORRECTED DRAWINGS ( as "replacement sheets") m	ust be submitted.	: :		
(a) ☐ including changes required by the Notice of Draftspe	erson's Patent Drawing Rev	riew ( PTO-948)	attached	
1) 🗌 hereto or 2) 🔲 to Paper No./Mail Date	_·			
(b) ☐ including changes required by the attached Examine Paper No./Mail Date	er's Amendment / Commen	t or in the Office	e action of	
Identifying indicia such as the application number (see 37 CFF each sheet. Replacement sheet(s) should be labeled as such in	R 1.84(c)) should be written on the header according to 37	n the drawings i CFR 1.121(d).	n the front (not the back) of	
6. DEPOSIT OF and/or INFORMATION about the department of attached Examiner's comment regarding REQUIREMEN	oosit of BIOLOGICAL MA	ATERIAL must BIOLOGICAL M	t be submitted. Note the MATERIAL.	
		:		
	· ·	:		
Attachment(s)				
1. Notice of References Cited (PTO-892)	5. Notice o	f Informal Paten	nt Application (PTO-152)	
2. Notice of Draftperson's Patent Drawing Review (PTO-948		v Summary (PT	O <del>-4</del> 13),	

U.S. Patent and Trademark Office PTOL-37 (Rev. 7-05)

Paper No./Mail Date \_\_\_\_\_

of Biological Material

3. Information Disclosure Statements (PTO-1449 or PTO/SB/08),

4. Examiner's Comment Regarding Requirement for Deposit

9. Dother\_

7. Examiner's Amendment/Comment

8. 

Examiner's Statement of Reasons for Allowance

Art Unit: 3739

## **DETAILED ACTION**

## **Reasons for Allowance**

- [01] The following is an examiner's statement of reasons for allowance.
- [02] Claim 4 is allowable because the Prior Art does not disclose a balloon supported by the header plate substantially surrounds the working lumen exit port and comprises a proximal balloon segment enclosing a proximal balloon segment inflation space and a distal balloon segment enclosing a distal balloon segment inflation space; and the inflation lumen of the access instrument body comprises a proximal balloon segment inflation lumen having an inflation lumen distal end opening into the proximal balloon segment inflation space and a distal balloon segment inflation lumen having an inflation lumen distal end opening into the distal balloon segment, whereby the proximal and distal balloon segments can be selectively inflated by inflation media introduced into the proximal or distal balloon segment inflation space.
- [03] Claim 31 is allowable because the Prior Art does not disclose a method which includes: surgically creating an incision into the anatomic space between the first and second anatomic surfaces; inserting the distal header through the incision with the first plate side disposed facing the first anatomic surface; advancing the distal header through the anatomic space between the first and second anatomic surfaces with the first plate side disposed facing the first anatomic surface; inflating the balloon to expand the anatomic space between the first and second anatomic surfaces; introducing one of a visualization instrument through the

Application/Control Number: 10/777,249

Art Unit: 3739

working lumen to visualize the anatomic space and the first anatomic surface to select a site of the first anatomic surface, a medical instrument through the working lumen for performing a medical procedure, an implantable medical device through the working lumen into the anatomic space, and a therapeutic drug or diagnostic agent through the working lumen into the anatomic space.

[04] Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

## Conclusion

- [05] Any inquiry concerning this communication or earlier communications from the examiner should be directed to Philip R. Smith whose telephone number is (571) 272 6087. The examiner can normally be reached between 9:00am and 5:00pm.
- [06] If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on (571) 272 4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.
- [07] Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only.

Art Unit: 3739

For more information about the PAIR system, see http://pair-direct.uspto.gov.

Should you have questions on access to the Private PAIR system, contact the

Electronic Business Center (EBC) at 866-217-9197 (toll-free).

John P. Leubecker Primary Examiner